Product Datasheet

PET Dose Adminstration System

Allows for safe drawing, dispensing and administration of PET isotopes.



Key Features

- Powder Coated Finish

Operation

1. Place the syringe into the BriTec PET syringe shield.
2. Load syringe from vial dispensing station.
3. Remove loaded syringe in shield and place into mobile pot using the appropriate plastic sleeve.
4. Ensure that the syringe is capped and that it fits into the pot vertically.
5. Place lead shielded caps on top and bottom.
6. Take to patient and, after locking the castors, gently release the clamp knob to rotate the pot through 80° to face the patient.
7. Tighten up the clamp knob, remove leaded caps and inject the patient.
8. Rotate the pot after use to the vertical position and dispose of the used syringe where appropriate.

Specifications

Code	Description	Shielding (mm)	External Dimensions (mm)
8BT-PETADMINSYS	**PET Dose Administration System**	12.5 min. (excludes syringe shield)	410 x 410 x 890

Bright Technologies Ltd
Bri-Tec House, 1 Holbrook Close
Holbrook Commerce Park
Sheffield, S20 3FJ

E-mail: info@britec.net
Tel: +44 (0)114 251 1215

www.britec.net

Version 1.0 May 2026